PR no: 2023-012

Department/Division: Operation/Water Quality

Colitag Specification:

General requirements: 100 pieces per box, Regeant pack inside sachet with easy tear, Brand as Colitag Water Test System

Physical Characteristics: Appearance: Homogenous powder, free flowing; Rehydrated: Colorless to pale straw; Weight of media: 1.2-1.4 g; pH: 6.6-7.0

Culture response: Media blank: None; Escherichia coli: Good, Fluorescence (MUG) + and Color Change (ONPG) +

Prepared by:	Checked by:	Approved by:
Granville T. Vosotros Water Utilities Management Officer B	Ma. Lourdes B. Rimas Quality Control Officer	Engr. Joselito A. Gillera Department Manager