



Calamba Water District

Lakeview Subdivision, Halang, Calamba, Laguna
Tel. Nos. 545-1614; 245-2086; 245-3180/3182
Tel./Fax: (049) 545-2863

PROJECT : SUPPLY & DELIVERY OF TOOLS, SHOP & GARAGE EQUIPMENT

SUBJECT : TECHNICAL SPECIFICATIONS

DATE : MARCH 7, 2019

1. COATING THICKNESS GAUGE

The device should have the following specifications:

Basic Model- Integral Probe	
Range	0 - 1500 μ m; (0-60mils)
Accuracy	\pm 1- 3% or \pm 2.5 μ m (\pm 0.1mil)
Resolution	0.1 μ m: 0 - 100 μ m; 1 μ m: 100-1500 μ m (0.01mil: 0-5 mils; 0.1mil: 5-60 mils)
Data Storage	5 last readings
Functions	Basic statistics
	Hi/ Low; Average; Number of Readings
	For Ferrous Substrates

The device shall include the following to its package.

- a. Calibration Foil Set
- b. Protective Case
- c. Instruction Manual

The device shall have two (2) years warranty guarantee.

The supplier shall held free training on proper handling, usage and maintenance of the unit.